

**CERTIFICATE OF TRACEABLE CALIBRATION**

No: CA052984

Page 1 of 28

CALIBRATION OF:

Sound Level Meter	2260	No: 2497376
Microphone:	4189	No: 2508721
Identification:		
Date of receipt:	2005-11-07	

CLIENT:

Vb Leasing D.o.o.
Zitnjak bb
Attn: Sani Sola, Vertriebsleiter
1000 Zagreb
Croatia

CALIBRATION CONDITIONS:

Preconditioning:	4 hours at 23 °C	
Environment conditions:	Air temperature:	23.0 °C ± 3°C
	Air pressure:	101.3 kPa ± 3 kPa
	Relative Humidity:	50.0 %RH ± 25 %RH

SPECIFICATIONS:

The Sound Level Meter has been calibrated in accordance with the requirements as specified in IEC60651 and IEC 60804.

PROCEDURE:

The measurements have been performed with the assistance of Brüel & Kjær Sound Level Meter Calibration System B&K 3630 with application software type 7763 and test collection 2260-4189-BZ7210-V2.0

RESULTS:

<input checked="" type="checkbox"/> Initial calibration	Calibration prior to repair/adjustment
<input type="checkbox"/> Recalibration without repair/adjustment	Calibration after repair/adjustment

The reported expanded uncertainty is based on the standard uncertainty multiplied by a coverage factor $k = 2$ providing a level of confidence of approximately 95 %. The uncertainty evaluation has been carried out in accordance with EA-4/02 from elements originating from the standards, calibration method, effect of environmental conditions and any short time contribution from the device under calibration.

Date of Calibration: 2005-11-07

Certificate issued: 2005-11-08

Kristian K. Nielsen

Calibration Technician

Peter Gaardsdal

Approved signatory: